



TS7500 OR Table with stationary column and Carbon Spine Surgical Tabletop



CARBON SPINE SURGICAL TABLETOP

OPTIMIZE SPINE SURGERY WORKFLOW

Enhanced Surgical Patient Access

Enables complex spine surgery with Hillrom's ergonomic and flexible Carbon Spine Surgical Tabletop.

- Optimal surgical site access due to low tabletop height.
- Supports a vast variety of individual patient anatomies due to flexible use of special positioning components and pads.
- Safe and ergonomic patient transfer handling due to lean and intuitive design.


The Carbon Spine Surgical Tabletop supports the natural anatomy of the patient for positive clinical outcomes.



SAFE, ERGONOMIC & FLEXIBLE

360° RADIOLUCENCY

Enhanced intraoperative imaging capabilities with up to 1640 mm (64.6") of radiolucent coverage in both standard or reverse patient orientation.



ADVANCED PATIENT POSITIONING

Flexible positioning of individual pads, components, and accessories for various patient anatomies.



GENERAL TO SPECIALTY

Transform your OR to accommodate multiple specialties with a single TS7500 Column and a wide range of tabletops.



TECHNICAL DETAILS

Material No.	# 2065423
Dimensions (L x W) Carbon Frame	2050 mm x 432 mm (80.7" x 17")
Height Range With a stationary column	620 mm - 1170 mm (24.4" - 46")
X-ray Area	Stationary column: 1640 mm x 432 mm (64.6" x 17") Mobile column: 1440 mm x 432 mm (56.7" x 17")
Longitudinal Slide	Stationary column: 800 mm (31.5") Mobile column: 600 mm (23.6")
Trendelenburg / Reverse Trendelenburg	± 20°
Tilt	± 15°
Maximum Patient Weight	225 kg (496 lbs)
Adjustment Speed of Longitudinal Slide	80 mm/s (3.1"/s)
Net Weight	Ca. 95 kg (ca. 209.4 lbs)

For more information please contact your local Hillrom representative.

hillrom.com

Hillrom
Amerika Building (7th Floor),
Hoogoorddreef 15
1101 BA Amsterdam, The Netherlands

TS7500 naming corresponds to the following CE registered products: TS 7500 - 4091000

The Carbon Spine Surgical Tabletop is intended to be used by clinicians and medically qualified personnel. This medical device is a regulated healthcare product which, pursuant to such regulation bears a CE mark. Hill-Rom recommends that you carefully read the detailed instructions for safe and proper use included in the documents accompanying the medical devices. The personnel of healthcare establishments are responsible for the proper use and maintenance of these medical devices. Hill-Rom reserves the right to make changes without notice in design, specifications and models. The only warranty Hill-Rom makes is the express written warranty extended on the sale or rental of its products. TRUMPF Medizin Systeme GmbH + Co. KG is a subsidiary of Hill-Rom Holdings, Inc.